

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\INVESTIGATORS\Buckner\PPG_Project1_MMem\T2_SPACE

TA: 2:31 PAT: 2 Voxel size: 1.2x1.2x1.2 mm Rel. SNR: 1.00 SIEMENS: tse_vfl

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Position	R3.0 A12.0 F18.0
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	12.50 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	144
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	1.20 mm
TR	2800 ms
TE	327 ms
Averages	1.0
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Reconstruction	Magnitude
Measurements	1
Multiple series	Off

Resolution

Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	6/8
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Integrated
Image Filter	Off
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On

Normalize	Off
B1 filter	Off
Raw filter	On
Intensity	Weak
Slope	25
Elliptical filter	Off

Geometry

Special sat.	None
Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain Atlas
Auto Coil Select	Off
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	R3.0 A3.0 H0.0
! Orientation	T > C-12.5
! Rotation	0.00 deg
! R >> L	216 mm
! A >> P	216 mm
! F >> H	141 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	Off
Dimension	3D
Bandwidth	651 Hz/Px

SIEMENS MAGNETOM TrioTim syngo MR B17

Flow comp.	No
Allowed delay	0 s
Echo spacing	3.42 ms
Adiabatic-mode	Off

Define	Echo trains
Turbo factor	109
Slice turbo factor	2
Echo trains per slice	1
Echo train duration	694
RF pulse type	Normal
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var